

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 02/24/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:45
INTERNAL STANDARD	VERIFIED	10:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:46
BLANK TEST	0.00	10:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:47
BLANK TEST	0.00	10:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:49

