

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 02/19/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:44
INTERNAL STANDARD	VERIFIED	18:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:45
BLANK TEST	0.00	18:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:46
BLANK TEST	0.00	18:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:47
BLANK TEST	0.00	18:48

