

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 02/29/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:36
INTERNAL STANDARD	VERIFIED	15:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:36
BLANK TEST	0.00	15:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:37
BLANK TEST	0.00	15:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:39
BLANK TEST	0.00	15:39

