

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 03/02/2020

OPERATOR NAME:  
CORBET W SULLIVAN  
Permit #: 45898  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:06
INTERNAL STANDARD	VERIFIED	08:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:06
BLANK TEST	0.00	08:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:09
BLANK TEST	0.00	08:10

*CS*  
*#623*  
*SOLUTION CHANGE*

