

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 02/29/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:18
INTERNAL STANDARD	VERIFIED	15:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:18
BLANK TEST	0.00	15:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:20
BLANK TEST	0.00	15:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:21
BLANK TEST	0.00	15:22

