

SOLUTION CHANGE

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 03/04/2020

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:50
INTERNAL STANDARD	VERIFIED	18:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:50
BLANK TEST	0.00	18:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:52
BLANK TEST	0.00	18:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:53
BLANK TEST	0.00	18:54

