

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 03/10/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:48
INTERNAL STANDARD	VERIFIED	23:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:51
BLANK TEST	0.00	23:52

