


**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Cambridge Village Police Department
Manufacturer: GUTH LABORATORIES Model: 12V500
Serial Number: MP4081
Date: January 13, 2020

Breath Simulator Temperature
34.00 degrees centigrade




Joseph Iavarone
Highway Safety Equipment Technician
Office of Public Safety

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Cambridge Village Police Department
Manufacturer: GUTH LABORATORIES
Model: 12V500
Serial Number: MP4081
Date: January 13, 2020

1. General appearance and condition of this device -----: OK
2. Check propeller rotation -----: OK
3. Check for air leaks -----: OK
4. Check for proper temperature -----:OK
5. Record any repairs made to this instrument and list services provided: **VERIFIED CALIBRATION**



Joseph Iavarone
Highway Safety Equipment Technician
Office of Public Safety

625 East Bunker Court
 Vernon Hills, Illinois 60061
 PH: 866-466-6225
 Fax: 847-327-2993
 www.innocalsolutions.com

NIST Traceable Calibration Report



Reference Number: 1371474
 PO Number: MC03222019

NY State Division of Criminal Justice Services
 80 South Swan St
 HSTU 3rd Floor
 Albany, NY 12210 United States

Manufacturer: Ertco	Calibration Date: 04/04/2019
Model Number: 139000-45	Calibration Due Date: 04/04/2020
Description: Temperature, Digital with Probe	Condition As Found: In Tolerance
Asset Number: 304001	Condition As Left: In Tolerance, No adjustment
Serial Number: 304001	
Procedure: DS Eutechnics 4400-EP642	

Remarks:

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Utilized

Asset No.	Manufacturer	Model No.	Description	Cal. Date	Due Date
CP144317	Fluke Corporation	1595A	Thermometer, Readout, Precision Thermometer	12/16/2017	12/31/2019
CP144386	Hart Scientific	5699	Temperature, Probe, SPRT	12/28/2017	10/31/2019
CP32033	Hart Scientific	5628	Temperature, PRT- Sensor	08/22/2017	08/31/2019
CP32034	Hart Scientific	5628	Temperature, PRT- Sensor	07/18/2018	01/31/2020
CP53595	Hart Scientific	5628	Temperature, PRT- Sensor	07/30/2018	01/31/2020

Calibration Data

FUNCTION TESTED	Nominal Value	As Found	Out of Tol	As Left	Out of Tol	CALIBRATION TOLERANCE
Temperature Accuracy	-15.000 °C	-15.00		Same		-15.050 to -14.950 °C [EMU 0.0098 °C][TUR 5.1:1]
	0.000 °C	0.01		Same		-0.030 to 0.030 °C [EMU 0.0072 °C][TUR 4.2:1]
	75.000 °C	74.99		Same		74.970 to 75.030 °C [EMU 0.011 °C][TUR 2.7:1]
	125.000 °C	124.98		Same		124.950 to 125.050 °C [EMU 0.015 °C][TUR 3.4:1]

Temperature: 22° C
Humidity: 24% RH
Rpt. No.: 1576270

Calibration Performed By:				Quality Reviewer:	
Sittinger, Jennifer	330	Metrologist	847-327-5317	Pietronico, Mike	04/04/2019
Name	ID #	Title	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, ISO 9002-94, and ISO 17025:2005. Guard Banding, if reported on this certificate, is applied at a Z-factor of 30% for test points with a test uncertainty ratio (TUR) below 4:1. In Tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.



New York State Division of Criminal Justice Services
SERVICE AUTHORIZATION FORM

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE.

SECTION I - AGENCY INFORMATION

Name of Agency Cambridge-Greenwich PD		Today's Date 1-13-2020
Address 56 North Park St		City, State, ZIP Cambridge NY 12816
Contact Person Officer Davis-Flynn		Contact Telephone 518-209-7141
Contact Telephone	Contact Facsimile	Contact Email

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer Guth Laboratories, Inc		Model Number 12V500
Instrument Serial Number 53132 MP4081		Antenna Serial Number (if applicable)
Date of Malfunction	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction SIM, Bottle, Bottle Holder, power cable, SIM Temp Cable.		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units.		
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein.		
Signature [Signature]		Date 01-13-20

DCJS USE ONLY

<u>EXPORT</u>	<u>Digital Therm.</u>	<u>Therm.#</u>	<u>APPROP-OK</u>
34.00°C	34.00°C		
34.00°C	34.00°C		
34.00°C	34.00°C	304001	AIRPAK
		4/4/20	check - OK