

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 03/18/2020

OPERATOR NAME:  
ARNOLD J RICHARD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:46
INTERNAL STANDARD	VERIFIED	04:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:46
BLANK TEST	0.00	04:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:48
BLANK TEST	0.00	04:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:49
BLANK TEST	0.00	04:50

