

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 03/09/2020

OPERATOR NAME:
 MATTHEW K JACKSON
 Permit #: 40078
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:49
INTERNAL STANDARD	VERIFIED	06:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53

