

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/23/2020

OPERATOR NAME:
ETHAN J MACURA
Permit #: 44080
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:54
INTERNAL STANDARD	VERIFIED	18:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:55
BLANK TEST	0.00	18:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:57
BLANK TEST	0.00	18:58

EM
#696

