

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 03/28/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:30
INTERNAL STANDARD	VERIFIED	07:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:31
BLANK TEST	0.00	07:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:32
BLANK TEST	0.00	07:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:33
BLANK TEST	0.00	07:34

