

SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 04/01/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:08
INTERNAL STANDARD	VERIFIED	15:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:08
BLANK TEST	0.00	15:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:09
BLANK TEST	0.00	15:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:11
BLANK TEST	0.00	15:11

