

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 03/30/2020

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:28
INTERNAL STANDARD	VERIFIED	21:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:28
BLANK TEST	0.00	21:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:29
BLANK TEST	0.00	21:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:31
BLANK TEST	0.00	21:31

