

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/30/2020

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:48
INTERNAL STANDARD	VERIFIED	13:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:48
BLANK TEST	0.00	13:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:49
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:51
BLANK TEST	0.00	13:52

Handwritten signature and number #691

