

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 04/18/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:08
INTERNAL STANDARD	VERIFIED	06:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:08
BLANK TEST	0.00	06:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:11
BLANK TEST	0.00	06:12

