

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

ate: 04/18/2020

OPERATOR NAME:
JEFFREY S DAIGLE
Permit #: 41298
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:21
INTERNAL STANDARD	VERIFIED	20:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:21
BLANK TEST	0.00	20:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:23
BLANK TEST	0.00	20:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:24
BLANK TEST	0.00	20:25

