

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 04/25/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:33
INTERNAL STANDARD	VERIFIED	07:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:36
BLANK TEST	0.00	07:36

