

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 04/29/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:31
INTERNAL STANDARD	VERIFIED	18:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:31
BLANK TEST	0.00	18:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:32
BLANK TEST	0.00	18:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:34
BLANK TEST	0.00	18:35

