

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 05/06/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:39
INTERNAL STANDARD	VERIFIED	01:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:39
BLANK TEST	0.00	01:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:42
BLANK TEST	0.00	01:43

