

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/27/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:09
INTERNAL STANDARD	VERIFIED	07:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:10
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:11
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:12
BLANK TEST	0.00	07:13

