

SOLUTION CHANGE

 GRANVILLE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 113306

Date: 05/04/2020

OPERATOR NAME:
 RYAN S PEDONE
 Permit #: 35878
 Operator Agency/Dept: GRANVILLE VILLAGE PD
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:53
INTERNAL STANDARD	VERIFIED	12:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:56
BLANK TEST	0.00	12:56

