

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 05/11/2020

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:59
INTERNAL STANDARD	VERIFIED	06:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:00
BLANK TEST	0.00	07:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:02
BLANK TEST	0.00	07:03

