

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 05/18/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:49
INTERNAL STANDARD	VERIFIED	06:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:52

