

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 05/04/2020

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:03
INTERNAL STANDARD	VERIFIED	22:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:06

