

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 04/06/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:41
INTERNAL STANDARD	VERIFIED	10:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:42
BLANK TEST	0.00	10:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:43
BLANK TEST	0.00	10:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:44
BLANK TEST	0.00	10:45

