

SUPERVISOR TEST

-----  
 FORT EDWARD  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 115708  
 -----

Date: 04/20/2020

OPERATOR NAME:  
 SGT DEAN E WATKINS  
 Permit #: 39523  
 Operator Agency/Dept: FORT EDWARD VILL. PD  
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:11
INTERNAL STANDARD	VERIFIED	11:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:12
BLANK TEST	0.00	11:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:13
BLANK TEST	0.00	11:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:14
BLANK TEST	0.00	11:15

