

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 05/18/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:58
INTERNAL STANDARD	VERIFIED	18:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:01
BLANK TEST	0.00	19:02

