

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 05/18/2020

OPERATOR NAME:
R M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:06
INTERNAL STANDARD	VERIFIED	07:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:06
BLANK TEST	0.00	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:07
BLANK TEST	0.00	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:09
BLANK TEST	0.00	07:09

