

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 05/27/2020

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:03
INTERNAL STANDARD	VERIFIED	19:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:03
BLANK TEST	0.00	19:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:05
BLANK TEST	0.00	19:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:06
BLANK TEST	0.00	19:07

