

SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 06/06/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:42
INTERNAL STANDARD	VERIFIED	06:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:42
BLANK TEST	0.00	06:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:44
BLANK TEST	0.00	06:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:45
BLANK TEST	0.00	06:46

