

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 06/21/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:11
INTERNAL STANDARD	VERIFIED	01:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:11
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:12
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:14
BLANK TEST	0.00	01:15

