

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 06/09/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:02
INTERNAL STANDARD	VERIFIED	00:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:06

