

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 06/15/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:42
INTERNAL STANDARD	VERIFIED	15:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:42
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:45
BLANK TEST	0.00	15:46

