

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 06/24/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:34
INTERNAL STANDARD	VERIFIED	18:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:35
BLANK TEST	0.00	18:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:36
BLANK TEST	0.00	18:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:37
BLANK TEST	0.00	18:38

