

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/15/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:25
INTERNAL STANDARD	VERIFIED	07:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:26
BLANK TEST	0.00	07:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:28
BLANK TEST	0.00	07:29

