

SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 06/01/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:27
INTERNAL STANDARD	VERIFIED	09:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:28
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31

