

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/28/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:00
INTERNAL STANDARD	VERIFIED	09:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:01
BLANK TEST	0.00	09:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:02
BLANK TEST	0.00	09:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:03
BLANK TEST	0.00	09:04

