

SOLUTION CHANGE

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/01/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:09
INTERNAL STANDARD	VERIFIED	12:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:10
BLANK TEST	0.00	12:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:11
BLANK TEST	0.00	12:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:12
BLANK TEST	0.00	12:13

