

SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 07/09/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:20
INTERNAL STANDARD	VERIFIED	06:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:20
BLANK TEST	0.00	06:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:22
BLANK TEST	0.00	06:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:23
BLANK TEST	0.00	06:24

