

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 07/18/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:58
INTERNAL STANDARD	VERIFIED	18:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:58
BLANK TEST	0.00	18:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01

