

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 07/11/2020

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:46
INTERNAL STANDARD	VERIFIED	22:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:47
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:48
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:49
BLANK TEST	0.00	22:50

