

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 06/08/2020

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:30
INTERNAL STANDARD	VERIFIED	20:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:31
BLANK TEST	0.00	20:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:32
BLANK TEST	0.00	20:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:34
BLANK TEST	0.00	20:35

[Handwritten Signature]
#691

