

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 06/15/2020

OPERATOR NAME:  
ETHAN J MACURA  
Permit #: 44080  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:32
INTERNAL STANDARD	VERIFIED	19:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:32
BLANK TEST	0.00	19:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:34
BLANK TEST	0.00	19:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:36
BLANK TEST	0.00	19:37

ETM  
#096

