

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 06/23/2020

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:08
INTERNAL STANDARD	VERIFIED	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:09
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:10
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:11
BLANK TEST	0.00	07:13

CM #821

