

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/14/2020

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:58
INTERNAL STANDARD	VERIFIED	16:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:58
BLANK TEST	0.00	16:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:00
BLANK TEST	0.00	17:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:01
BLANK TEST	0.00	17:02

Hurlburt
#257

