

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 05/26/2020

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:52
INTERNAL STANDARD	VERIFIED	16:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:53
BLANK TEST	0.00	16:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:54
BLANK TEST	0.00	16:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:55
BLANK TEST	0.00	16:56

