

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 06/01/2020

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:39
INTERNAL STANDARD	VERIFIED	21:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:41
BLANK TEST	0.00	21:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43

