

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 07/13/2020

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:04
INTERNAL STANDARD	VERIFIED	03:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:05
BLANK TEST	0.00	03:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:07
BLANK TEST	0.00	03:07

