

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 07/18/2020

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:46
INTERNAL STANDARD	VERIFIED	06:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:47
BLANK TEST	0.00	06:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50

